



DR. BABASAHEB AMBEDKAR TECHNOLOGICAL UNIVERSITY, LONERE

At. Po. Lonere, Tal. Mangaon, Dist. Raigad 402 103 MS (www.dbatu.ac.in)

INVITATION FOR QUOTATION

07 DEC 2020

Our Ref No.: DBATU/Civil Engg. /Env Lab/ /2020/ 1787

Date: 25/11/2020

Quotation for Supply of Environment Lab Equipment's

Due On: 16/12/2020

Date of Opening: 17/12/2020

Time: 3.00 pm

To,

Sub: QUOTATION FOR THE SUPPLY OF Environment Lab Equipment's

Dear Sir,

Your quotation for the items listed overleaf, may please be submitted to the under signed, so as to reach this Office not later than 16/12/2020. While submitting your quotation, the following procedure may please be observed and other points borne in mind.

01. The maker's name must be specified.
02. The "Terms and Conditions" for supply and delivery of stores, should be clearly indicated in the quotation, stating whether rates are, inclusive of all taxes, Packing and forwarding charges Freight charges, etc. or not, however rates offered as including all taxes will be more welcome.
03. If packing and forwarding charges are to be charged separately, it should be so clearly stated in your quotation.
04. Please mention clearly whether consignment would be Ex-Godown, Ex-Shop, of F.O.R. dispatching stations. Preferably terms offered as "Delivery of consignment of stores, on F.O.R."
05. Envelope should be clearly marked with our reference No and date of this quotation. It should also be superscripted as per the format given above.
06. The quotation would be opened as per date and timing given above, if desired by you, you may depute an authorized representative with a letter of authority to be present at the time of opening of the quotation at this Office on the aforesaid day, date & time.
07. Your quotation must be valid for a minimum 30 (Thirty) working days from the date of it's opening.
08. Quotation received after the date of opening may not be taken into consideration.
09. Items tendered should confirm to the specification shown in the attached list when and where, full or no specifications are indicated against items in the list. Kindly furnish your full specification in accordance with accepted standards against each item tendered. Where reference to catalogue is made, the relevant catalogues/ Pamphlets/ Literature should accompany the quotation.
10. Your quotation should be for all new items and not for second hand and should be submitted as per standard format of Quotation Submission attached herewith.
11. Please state whether items will be available Ex-Stock. If not the minimum period for delivery, or for supplying the items or stores.
12. It should be clearly stated whether **GST, Insurance, Freight or packing** and forwarding charges, or any other taxes and duties, etc. leviable.
13. It would be appreciated if illustrated catalogues/Literature etc is furnished with the quotation.

14. Expression to as "Complete with standard Equipment" complete with standard accessories "Equipment to" As good as should be avoided. If at all their use is unavoidable then it should be very specifically indicated as to what exactly they mean and what exactly would be supplied under them. Any ambiguity or vagueness should be avoided.
15. For convenience, kindly adopt while quoting the same serial Nos. as given in the list detailed below.

Thanking You.

Yours faithfully,


Registrar

Dr. Babasaheb Ambedkar Technological
University,

List of Items

Sr. No	Description / Specification	Make	Approx. Qty.
01	Technical Specifications of Environment Lab Equipment's are attached as a separate sheet	-	-

Detailed Technical Specifications (Environmental Engg. Laboratory Equipment's)

Sr. No	Name of Equipment's	Specification	Qty
1	pH Meter	<ul style="list-style-type: none"> • Range: 0.00 to 14.00 pH • Resolution: 0.01 pH • Accuracy: 0.01 pH \pm 1 digit • Dimensions: 205mm x 65mm x 130 mm • Table top with standard accessories 	1
2	Turbidity Meter cum Nephelometer	<ul style="list-style-type: none"> • Range: 0-1000 NTU • Resolution: 0.1 NTU in lowest • Accuracy: \pm 3% FS \pm 1 Digit • Detector: Photodiode/ Photocell • Calibration: manual. 	1
3	Conductivity Meter	<ul style="list-style-type: none"> • Range: 0- 20μs, 200 μs, 2 μs, and 200 ms/cms • Accuracy: \pm 1% F.S. in all ranges • Dimensions: 205mm x 65mm x 130 mm • Table top with standard accessories 	1
4	B. O. D. Incubator	<ul style="list-style-type: none"> • Double walled with inner chamber of stainless steel and outer wall of M. S. Sheet • Adjustable shelves • Capacity: 6 Cu. Ft. (170 Litres) • Inner Dimensions: 50x50x70 cm • Temp Range: 5° to 60° C • Accuracy: \pm 0.5° C 	1
5	Digital Analytical Balance	<ul style="list-style-type: none"> • Capacity: 300gm • Readability: 0.01 g • Repeatability: \pm 0.01g • Linearity: \pm 0.02 g • Pan size (mm): 128 dia. 	1
6	COD Digestion Apparatus	<ul style="list-style-type: none"> • Temp Range: above amb. to 180°C • Resolution: 1°C • Timer: selectable • Dimensions: 500 mm x 270 mm x 210 mm (w x d x h) • Hole size: 40 mm dia. and 80 mm depth • With accessories: glass tube, air condenser, tube stand and condenser stand • Glass tubes: 38mm dia. 6 nos' (3X2 rows) 	1
7	TDS Meter	<ul style="list-style-type: none"> • Table model • Ranges: 5 ranges 0-1000 ppt Mohs/cm • Temp. Compensation: manual 0 -50° C • Measuring frequency: 1000 Hz • Accuracy: + 1% FSI • Dimensions: 76 X 275 X 175 mm 	1
8	Dissolved Oxygen Meter	<ul style="list-style-type: none"> • Range: D.O: 0 to 20 ppm • Resolution: D.O.: 0.1 ppm • Accuracy: 0.1 ppm \pm 1 count • Temp. compensation: 0° C to 50° C manual 	1
9	Colorimeter	<ul style="list-style-type: none"> • Range: 400 - 700 nm • Output: OD (0 to 1.99) • Resolution: OD: 0.01 • Accuracy: 0.5 and FSI • Min. volume: 1 ml • With battery backup 	1
10	Jar Test Apparatus	<ul style="list-style-type: none"> • Range: variable speed up to 50 - 100 RPM with speed regulator • Stainless steel paddles • Stirrer: 6 	1

AAD arge

[Signature]

02/12

Format of Quotation Submission
(In Letter Head of Supplier with seal)

To, _____

Date: _____

Sr. No.	Description of goods (with full Specifications)	Qty.	Unit	Quoted Unit rate in Rs. (Including Ex Factory price, excise duty, packing and forwarding, transportation, insurance, other local costs incidental to delivery and warranty/ guaranty commitments)	Total Price (A)	GST and other taxes payable	
						In %	In figures (B)
Total Cost							

Gross Total Cost (A+B): Rs. _____

We agree to supply the above goods in accordance with the technical specifications for a total contract price of Rs. _____ (Amount in figures) (Rupees _____ amount in words) within the period specified in the Invitation for Quotations.

We confirm that the normal commercial warranty/ guarantee of _____ months shall apply to the offered items and we also confirm to agree with terms and conditions as mentioned in the Invitation Letter.

We hereby certify that we have taken steps to ensure that no person acting for us or on our behalf will engage in bribery.

Signature of Supplier

Name: _____

Address: _____

Contact No: _____